

Certificate



Certificate no.

TA 50053140 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Test report no.: ZTW1-BHU 11002852 001

Client Reference: 93-0685/PSE

Tested to:

EN 60601-1:1990+A1+A2+A13
IEC 60601-1:1988+A1+A2

Certified Product: (Open frame Power Supply)

License Fee - Units

Type Designation : MPx-120y (MW)
x stands for : S, D, T, Q
y stands for : -3.3, -5, -12, -15, -24, -48,
A, B, C, D, E
Rated Voltage : AC 100-240V, 50/60 Hz
Rated Current : 2.9-1.3A
Output Voltages : see constructional dataform
Output Currents : see constructional dataform
max. Ambient Temperature: 40°C
Protection Class : I

8
1
1




Remarks: Primary-and secondary circuits are separated according to method 4 of clause 17.g.

10

Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signatures

Date of Issue

(day/mo/yr)

26/11/2004

Dipl.-Ing. R. Scheller

Dipl.-Ing. A. Klinker

Certificate



Certificate no.

TA 50053140 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Test report no.: ZTW1-BHU 11002852 003

Client Reference: 0805006002-003/LDI

Tested to: EN 60601-1:1990+A1+A2+A13
IEC 60601-1:1988+A1+A2

Certified Product: (Open Frame Power Supply)

License Fee - Units

as page 01

Change


x stands for : S, D, T, Q

y stands for : -3.3, -5, -12, -15, -24, -48,
A, B, C, D, E, BC88


Output Voltage : see Appendix

Output Current : see Appendix

Appendix: 1



Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signature

Dipl.-Ing. F. Stöelzel

Date of Issue

(day/mo/yr)

16/05/2008