

Certificate



Certificate no.

TA 50079957 01

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Test report no.: ZTW1-BHU 10015994 001

Client Reference: 0512068001/LDI

Tested to:

EN 60601-1:1990+A1+A2+A13
IEC 60601-1:1988+A1+A2


Certified Product: (Switching Power Supply)

License Fee - Units

Type Designation : PM-15-X (MW)
X stands for : 3.3, 5, 12, 15, 24
for different output voltage
Rated Voltage : AC 100-240V, 50/60Hz
Rated Current : 0.35A-0.2A
Output Voltage : DC 3.3V, 5.0V, 12.0V, 15.0V, 24.0V
Output Current : 3.5A, 3.0A, 1.25A, 1.0A, 0.63A
max. Ambient Temperature: 40°C
Protection Class : II


8
1

Remark: Primary and secondary circuits are separated according to method 4 of clause 17.g.



Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.

Signatures

Date of Issue

(day/mo/yr)
24/02/2006

Dipl.-Ing. R. Scheller

Dipl.-Ing. A. Klinker

Certificate



Certificate no.

TA 50079957 02

License Holder:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Manufacturing Plant:

Mean Well Enterprises Co., Ltd.
No. 28, Wu-Chuan 3rd Road
Wu Ku Ind. Park, Taipei Hsien 248
Taiwan

Test report no.: ZTW1-BHU 10015994 001

Client Reference: 0512068001/LDI

Tested to:

EN 60601-1:1990+A1+A2+A13
IEC 60601-1:1988+A1+A2

Certified Product: (Switching Power Supply)

License Fee - Units

as page 01

Addition

Type Designation : NFM-15-X (MW)

1

X stands for : 3.3, 5, 12, 15, 24
for different output voltage


1

Appendix: 1



2

Licensed Test mark:



TÜV Rheinland Taiwan Ltd.


Signatures

Date of Issue

(day/mo/yr)

24/02/2006


Dipl.-Ing. R. Scheller


Dipl.-Ing. A. Klinker