

Certificate



Certificate no.

TA 50367142 01

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 50045257 001

Client Reference: FL_02102015_RPS-400

Tested to:

IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

Type Designation	: RPS-400-x (MW)	10
x stands for	: 12, 15, 18, 24, 27, 36 or 48	1
Rated Voltage	: AC 100-240V, 50/60Hz	
Rated Current	: 4.8-2.1A	
Outputs	: see constructional dataform	
max. Power Output	: see constructional dataform	
max. Ambient Temperature	: see constructional dataform	
max. Operational Altitude	: 4000m	
Protection Class	: I	

Remarks: 1) The risk management requirements of the standard were not addressed;
2) This certificate is valid for maximum 5 years.

11

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue

(day/mo/yr)

15/12/2016

Certificate



Certificate no.

TA 50367142 02

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Test report no.: ZTW1- 50045257 001

Client Reference: FL_02102015_RPS-400

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Addition

Type Designation	: RPS-400-x-C (MW),	1
	RPS-400-x-SF (MW),	1
	RPS-400-x-TF (MW)	1
x stands for	: 12, 15, 18, 24, 27, 36 or 48	1
max. Operational Altitude	: 4000m	

This certificate is valid for maximum 5 years.

4


Appendix: 1

Licensed Test mark:



TÜV Rheinland Taiwan Ltd
Signature

W. Hsu
Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

15/12/2016

Certificate



Certificate no.

TA 50367142 03

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

SuZhou MEAN WELL Technology
Co., Ltd.
No. 77, Jian-min Road,
Dong-qiao, Pan-yang Ind. Park,
Huang-dai Town, Xiang-cheng District
Suzhou, Jiangsu 215152
P.R. China

Test report no.: ZTW1- 50045257 001

Client Reference: FL_02102015_RPS-400

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

15/12/2016

Certificate



Certificate no.

TA 50367142 04

License Holder:

MEAN WELL Enterprises Co., Ltd.
No. 28, Wuquan 3rd Rd.,
Wugu Dist., New Taipei City 248
Taiwan

Manufacturing Plant:

MEAN WELL (GUANGZHOU) ELECTRONICS
CO., LTD HUADU BRANCH
No. 11 Jingu South Road,
Huadong Town,
Huadu District, Guangzhou 510890
P.R. China

Test report no.: ZTW1- 50045257 001

Client Reference: FL_02102015_RPS-400

Tested to: IEC 60601-1:2005+A1
EN 60601-1:2006+A11+A1+A12

Certified Product: (Switching Mode Power Supply)

License Fee - Units

as page 01

Addition

Factory : see above

This certificate is valid for maximum 5 years.

Appendix: 1


Licensed Test mark:



TÜV Rheinland Taiwan Ltd

Signature

Dipl.-Ing. W. Hsu



Date of Issue
(day/mo/yr)

15/12/2016